



Unit 9, Kingsthorpe Business Centre, Studland Road, Kingsthorpe, Northampton NN2 6NE  
Tel: +44 (0)1604 777700 Fax: +44 (0)1604 721844 Email: admin@bainternational.com

## *Certificate of Annual Service*

Statim Cassette Autoclave  
Serial Number: 320717G00018

In order to maintain a high level of performance from your equipment, BA International advocates regular service and maintenance protocols in accordance with manufacturers recommendations. This is to certify that on the 5<sup>th</sup> May 2023 a SciCan approved engineer inspected and serviced the above Statim Cassette Autoclave in-line with manufacturer's annual maintenance schedule. Following service, a series of functionality tests were carried out to ensure the unit is functioning within manufacturer specifications.

Functionality tests include:

- **Water Quality Sensor:** Ensure message "Water Quality Unacceptable" appears on screen when water is used that has a reading greater than 5 parts per million.
- **Pump Delivery:** Check pump delivery is within acceptable values.
- **Calibration Check:** Check that the theoretical and actual chamber temperatures are within acceptable values during hold phase.
- **Safety Checks:** Ensure cassette cannot be removed during a cycle.  
Check cassette for leaks.
- **Pipe work & Connections:** Check internal Pipe work and fittings for leaks.

**NB** This is not a validation certificate as required by decontamination guidance document HTM01-05.

Your next service is due 1 year from the above date.

Should you require any additional information regarding any of our products or services please do not hesitate to contact our Technical Services Department on 01604 777700, where a member of our Technical Support Team will endeavour to assist wherever possible.



SciCan Approved Service Centre

**SciCan**